

Bestimmung sehr niedriger Alkoholwerte (wässrige Standards)

Geräte: SF2010; SF 2030; Chromstar 6.4; manueller Injektor RH8125, jeweils 5µl-injektion

% Alkohol in H ₂ O	Peakfläche	Peakhöhe	% Alkohol in H ₂ O
0,01	4210	0,386	0,01
0,05	25930	2,035	0,05
0,1	52869	3,935	0,1
0,5	280391	20,524	0,5
1	553214	40,045	1



